

СООБЩЕНИЯ  
ОБЪЕДИНЕННОГО  
ИНСТИТУТА  
ЯДЕРНЫХ  
ИССЛЕДОВАНИЙ

Дубна

ff

CERN LIBRARIES, GENEVA



SCAN-0003041

E16-99-158

F.Spurný\*, B.Vlcek\*, V.P.Bamblevski, G.N.Timoshenko

SPECTRA OF THE LINEAR ENERGY TRANSFER  
MEASURED WITH A TRACK ETCH SPECTROMETER  
IN THE BEAM OF 1 GeV PROTONS  
AND THE CONTRIBUTION OF SECONDARY  
CHARGED PARTICLES TO THE DOSE

\*Department of Radiation Dosimetry, Nuclear Physics Institute,  
Czech Academy of Sciences, Prague, Czech Republic

1999

## I. INTRODUCTION

A spectrometer of the linear energy transfer (LET) based on the chemically etched polyallyldiglycolcarbonate (PADC) nuclear track detector (NTD) has been recently developed /1-3/. It has already been used to determine LET spectra in several radiation beams and fields /4-8/.

This report presents, discusses and analyzes the LET spectra established in the detectors irradiated by the beam of 1 GeV protons. The goal of experiments was to enlarge the studies already performed with the protons of energies of few hundred MeV /5-7/ and, also, to obtain new data on the dosimetric characteristics of beams supposed to be used for the programme of the «accelerator driven transmutation technologies» (ADTT) /9/.

## 2. EXPERIMENT

### 2.1. Irradiation

The irradiation has been carried out by the beam of 1 GeV protons of the LHE synchrotron. The monitoring of irradiation conditions has been ensured through the beam ionization chamber measurements. The proton flux during the irradiation was about  $1.65 \times 10^8 \text{ cm}^{-2}$ , that corresponded to the tissue kerma value about 50 mGy. The particle fluences have been determined by the sets of activation detectors (reactions  $^{12}\text{C} \rightarrow ^{11}\text{C}$ ;  $^{27}\text{Al} \rightarrow ^{24}\text{Na}$  or  $^{18}\text{F}$ ), placed perpendicularly to the beam simultaneously with the NTD. In total, five sets of detectors have been exposed.

### 2.2. LET Spectrometer Based on Chemically Etched PADC NTD

Polyallyldiglycolcarbonate available from Pershore Moulding, England (curing time 32 hours, thickness 0.5 mm) has been used. The detector samples have been etched in a 5N NaOH at 70°C after irradiation. Each sample was irradiated before etching in a corner with  $^{252}\text{Cf}$  fission fragments, in another one with  $^{241}\text{Am}$  alpha particles to check exact etching conditions and to determine the bulk etching. The optimal condition was corresponded to 18 hours of etching (one-side removed layer about 17  $\mu\text{m}$ ).

To determine LET-value of a particle, the etch rate ratio  $V = V_T/V_B$  (where  $V_B$  is bulk etching rate and  $V_T$  is track etching rate) has been primarily established through the determination of track parameters. The track parameters have been measured by means of an automatic optical image analyzer LUCIA II based on a Leitz microscope /1-3/. The value of  $V$  was calculated by at least two ways of track parameter combination, the final optimization is performed

through the comparison of the removed layer thickness recalculated from V-value and directly measured through fission fragment track diameter. The obtained V-spectra were corrected for the critical angle of the registration and transformed to LET spectra on the basis of the heavy charged particles calibration checked through high dose electron irradiation. The spectrometer permits to estimation LET of a track between 100 and 7000 MeV.cm<sup>2</sup>.g<sup>-1</sup> in tissue.

Two sheets of PADC NTD 30x30 mm<sup>2</sup> covered by one half with 2 mm of polyethylene (PE) represent an irradiation set. It is always irradiated being sealed in an aluminized foil. The sheets combination gives the possibility to have four radiator positions: bare, under PE, under the first of PADC sheet (bare-CR) and under PADC covered with PE (CR-PE).

### **3. RESULTS AND DISCUSSION**

#### **3.1. LET spectra**

The LET spectrometer permits to establish the distributions in LET of both the absorbed dose D and the equivalent dose H. For these purposes quality factors from both recent ICRP Recommendations /10,11/ have been used. First, we have observed that there is a minimal influence of the radiator on both absolute values of linear energy transfer as well as their spectral shape. Typical examples of the obtained results are for the absorbed dose and the equivalent dose with ICRP 60 quality factors presented in Figures 1 and 2. One can see there that the both spectra as well as absolute values of differential dose quantities are quite similar for all four radiators. A good agreement has been also observed for the exposed five detector sets.

More information on the contribution of secondary charged particles with the different LET values to the total LET value can be obtained from so-called microdosimetric distributions L\*D(L) and L\*H(L) respectively. These distributions are presented in Figure 3, for the absorbed dose as well as for the equivalent doses with both sets of quality factors.

One can see that all three spectra are peaked in the region about 2000 MeV.cm<sup>2</sup>.g<sup>-1</sup>, the average values of LET with respect to the energy delivery can be estimated to about 1600 MeV.cm<sup>2</sup>.g<sup>-1</sup> for the absorbed dose and the equivalent dose (ICRP 21). It is a little less in the case of the equivalent dose with ICRP 60 quality factors.

#### **3.2. Integral dosimetry and microdosimetry characteristics**

The LET absorbed dose and equivalent dose distributions enable us to calculate the integral values of the absorbed dose D and equivalent dose H,

corresponding to the secondary particle tracks which are revealed. These integral values are obtained as:

$$D = \int (dN / dL) \cdot L \cdot dL ; \quad (1)$$

$$H = \int (dN / dL) \cdot L \cdot Q(L) \cdot dL ; \quad (2)$$

where  $dN/dL$  is the number of tracks in a LET interval;  $L$  is the value of LET; and  $Q(L)$  is the quality factor corresponding to the value of  $L$ .

The integral characteristics obtained from the LET distributions through the equations (1) and (2) are presented in Table 1. As discussed in previous papers /7/, the values presented in Table 1 could represent the total contribution of high LET particles to the tissue dose characteristics in the beam of high energy protons. One can see that these contributions increase regularly with the proton energy, going for the tissue kerma and 1 GeV protons up to about 6.8 %

Table 1: Total dose and dose equivalents due to track forming secondary particles created in the PADC LET spectrometer under high energy proton irradiation (primary proton tissue kerma = 1 Gy).

Quantity; unit	$E_p$ , GeV			
	1.0	0.2	0.16	0.085
D, mGy	68.2±6.2	27.2±2.1	22.5±2.0	12.2±0.6
H (ICRP21), mSv	782±80	277±25	187±21	140±6
H (ICRP60), mSv	944±92	327±28	225±24	175±7

Of course, the quality factors defined for radiation protection purposes are not appropriate to appreciate, for example, relative biological efficiency in the radiotherapy. Nevertheless, the values of obtained «equivalent doses» are high enough (up to 90 % of tissue kerma due to ionization losses for 1 GeV protons) to be considered in the relation to possible differences in the efficiency of high energy protons as compared with photons and/or electrons with comparable microdosimetric characteristics for primary radiation. It is well known, for example, that the relative radiotherapy biological efficiency of few hundred MeV protons is estimated up to about 1.2 /12/. Nevertheless at radiobiological irradiation by a high energy particle beam even with thin biological samples it is very important to know the secondary charged particle contributions to the doses. The described technique of experimental estimation of LET –spectra in solid matter (the composition of CR-39 detector is tissue-equivalent roughly) allows one to define more exactly dosimetric characteristics and corresponding biological efficiencies at radiobiological experiments.

Work partially supported through the EC Project FI3P-CT92-0026 and the grant No. 3048606 of the IGA of the AS CR and the grant No. 202/0151/99 of the GACR.




Fig. 1. LET spectra of the absorbed dose due to the secondary particles in the 1 GeV proton beam




Fig. 2. LET spectra of the equivalent dose (ICRP60) due to the secondary particles in the 1 GeV proton beam




Fig.3. Microdosimetric distributions of the absorbed and equivalent doses due to the secondary particles in the 1 GeV proton beam

## REFERENCES

1. J. Charvát: „LET spectrometry with polymer track etched detectors“. PhD Thesis, Prague 1986, pp. 1-111
2. F. Spurný, J. Bednár, L. Johansson, A. Säterberg. “LET spectra of secondary particles in CR39 track etch detectors“. *Radiation Measurements* 26, (1996), p. 645-649
3. F. Spurný, J. Bednár, K. Turek. “Spectrometry of linear energy transfer with a track etch detectors“. *Radiation Measurements* 28, (1997), p. 515-518
4. F. Spurný, L. Johansson, A. Säterberg, J. Bednár, K. Turek. “Contribution of Secondary Heavy Particles to the Absorbed Dose from High Energy Photon Beams“. *Phys. Med. Biol.* 41, (1996), p. 2643-2656
5. V. E. Dudkin, F. Spurný. “Contribution of Secondary High LET particles to the Energy Transfer from High Energy Protons“. *Proc. IRPA9 Congress, Vienna 1996, vol. 2., p. 250-253*
6. F. Spurný, J. Bednár, K. Turek, V. P. Bamblevski. “Spectrometry of the LET in High Energy Reference Fields.“ *Bezp. jaderné energie* 5 (43), (1997), p. 354-358
7. F. Spurný, J. Bednár, K. Turek, J.-F. Bottollier-Depois, V. E. Dudkin. “Secondary Particles Contribution to the Dose Characteristics in High Energy Particle Beams and Fields“. *Radiation Measurements* 28, (1997), p. 505-513
8. J.-F. Bottollier-Depois, F. Spurný, L. Plawinski, I. Votocková, J. Bednár, M. Viso and A. Labarthe. “Dosimetry during the First IBIS Facility Flight“. *Adv. Space Research* 22, (1998), p. 517-520
9. Accelerator Driven TT;
10. “Recommendations of the International Commission on Radiological Protection“. ICRP Publication 26, *Annals of ICRP*, 1, No. 3, (1977)
11. “Recommendations of the International Commission on Radiological Protection“. ICRP Publication 60, *Annals of ICRP*, 21, No. 1-3, (1990)
12. A. Wambersie. “Dose Specification in Heavy Particle Therapy“. *Radiat. Prot. Dosim.* 70, (1997),

Received by Publishing Department  
on May 31, 1999.

**The Publishing Department  
of the Joint Institute for Nuclear Research  
offers you to acquire the following books:**

Index	Title
E1,2-97-79	Proceedings of the XIII International Seminar on High Energy Physics Problems. Relativistic Nuclear Physics and Quantum Chromodynamics. Dubna, 1996 (2 volumes, 364 p. and 370 p., in English)
D5,11-97-112	Proceedings of the 9th International Conference «Computational Modelling and Computing in Physics». Dubna, 1996 (378 p., in English)
E2-97-213	Proceedings of the V International Seminar on Interaction of Neutron with Nuclei «Neutron Spectroscopy, Nuclear Structure, Related Topics». Dubna, 1997 (446 p., in English)
E2,4-97-263	Proceedings of the Third International Conference «Renormalization Group'96». Dubna, 1996 (436 p., in English)
E10-97-272	Proceedings of the Data Acquisition Systems of Neutron Experimental Facilities (DANEF'97). Dubna, 1997 (325 p., in English)
D19-97-284	Proceedings of the International Symposium «Problems of Biochemistry. Radiation and Space Biology». Dubna, 1997 (2 volumes 284 p. and 405 p., in Russian and English)
E2-97-413	Proceedings of the VII Workshop on High Energy Spin Physics (SPIN'97). Dubna, 1997 (398 p., in English)
E15-98-57	Proceedings of the III International Workshop «Hyperfine Structure and Nuclear Moments of Exotic Nuclei by Laser Spectroscopy». Poznan, 1997 (200 p., in English)
D13-98-66	Proceedings of the XVII International Symposium on Nuclear Electronics. Varna, 1997 (242 p., in English and Russian)
E1,2-98-154	Proceedings of the XIII International Seminar on High Energy Physics Problems «Relativistic Nuclear Physics and Quantum Chromodynamics». Dubna, 1996 (2 volumes 300 p. and 282 p., in English)
E3,14-98-168	Proceedings of the German–Russian User Meeting «Condensed Matter Physics with Neutrons at IBR-2». Dubna, 1998 (126 p., in English)
E3-98-202	Proceedings of the VI International Seminar on Interaction of Neutrons with Nuclei Neutron Spectroscopy, Nuclear Structure, Related Topics. Dubna, 1998 (352 p., in English)
D1,2-98-215	Proceedings of the International Workshop «Relativistic Nuclear Physics: from MeV to TeV». Dubna, 1998 (384 p., in English and Russian)
E2-98-254	Proceedings of the International Workshop Hadronic Atoms and Positronium in the Standard Model. Dubna, 1998 (260 p., in English)



Index	Title
D9,11-98-273	Proceedings of the 4th International Workshop «Beam Dynamics and Optimization». Dubna, 1997 (162 p., in English and Russian)
E17,19-98-305	Proceedings of the International Workshop on Deuteration of Biological Molecules for Structural and Dynamic Studies. Applications to Neutron Scattering and NMR. Dubna, 1998 (100 p., in English)
E1,2-98-307	Proceedings of the International School–Seminar «Actual Problems of Particle Physics». Gomel, 1997, Belarus (2 volumes 304 p. and 220 p., in English)
E2-98-372	Proceedings of the III International Workshop «Classical and Quantum Integrable Systems». Erevan, 1998 (200 p., in English)
E9-99-26	Proceedings of the International Conference on High Energy Accelerators. Dubna, 1998 (432 p., in English)
E2-99-35	Proceedings of the XI International Conference on Problems of Quantum Field Theory. Dubna, 1998 (508 p., in English)
E5-99-38	Proceedings of the International Workshop «Self-Similar Systems». Dubna, 1998 (404 p., in English)
E9-99-92	Proceedings of the XVII International Conference on High Energy Accelerators HEACC'98. Dubna, 1998 (435 p., in English)

Please apply to the Publishing Department of the Joint Institute for Nuclear Research for extra information. Our address is:

Publishing Department  
 Joint Institute for Nuclear Research  
 Dubna, Moscow Region  
 141980 Russia  
 E-mail: [publish@pds.jinr.dubna.su](mailto:publish@pds.jinr.dubna.su).

Спурны Ф. и др.

E16-99-158

Измерение спектров ЛПЭ в пучке протонов с энергией 1 ГэВ с помощью трекового спектрометра и вклад в дозу от вторичных заряженных частиц

Для оценки линейной передачи энергии (ЛПЭ) в пучке протонов с энергией 1 ГэВ использовался трековый спектрометр на основе детектора CR-39. Показано, что спектры ЛПЭ вторичных заряженных частиц в диапазоне от 100 до 7000 МэВ см<sup>2</sup>г<sup>-1</sup> не зависят от используемого радиатора. Средние значения фактора качества вторичных заряженных частиц в указанном диапазоне ЛПЭ составили 11,6 при использовании данных Публикации 26 МКРЗ и 14,0 при использовании данных Публикации 60 МКРЗ.

Полученные спектры позволили оценить вклад вторичных заряженных частиц в дозиметрические величины. Установлено, что их вклад в полную поглощенную дозу протонов составляет около 7%, а в эквивалентную дозу — до 90%.

Работа выполнена в Отделении радиационных и радиобиологических исследований ОИЯИ.

Сообщение Объединенного института ядерных исследований. Дубна, 1999

Spurný F. et al.

E16-99-158

Spectra of the Linear Energy Transfer Measured with a Track Etch Spectrometer in the Beam of 1 GeV Protons and the Contribution of Secondary Charged Particles to the Dose

A spectrometer of the linear energy transfer (LET) on the base of CR-39 detector was used to establish the spectra of LET in the beam of protons with the primary energy of 1 GeV. It was found out that the LET spectra of secondary charged particles between 100 and 7000 MeV cm<sup>2</sup>g<sup>-1</sup> do not depend on the radiator. The average quality factors for the LET region mentioned were obtained about 11.6 with ICRP 26 quality factors and about 14.0 with ICRP 60 quality factors.

The spectra obtained permitted to calculate the contributions of these secondary charged particles to the dosimetric quantities. It was observed that these contributions were about 7.0% for the total absorbed dose of protons and close to 90% in the case of the equivalent doses. It is more than it was found out for few hundred MeV protons.

The investigation has been performed at the Division of Radiation and Radiobiological Research, JINR.

Communication of the Joint Institute for Nuclear Research. Dubna, 1999