

Supplementary file

Table S1. Accuracy data for all samples in 3 different groups.

Implant ID	Printed Group (Segmented vs Printed Tooth) n=14			Segmentation Group (Segmented vs Actual Tooth) n=8*			Overall Group (Printed vs Actual Tooth) n=8*		
	Accuracy (%)	Min(%)	Max(%)	Accuracy (%)	Min(%)	Max(%)	Accuracy (%)	Min(%)	Max(%)
AL	83.497	-0.1126	0.2848	66.020	-0.3359	0.7368	73.190	-0.7867	0.2374
AR	90.918	-0.0937	0.3433	68.200	-0.1912	0.421	69.750	-0.4315	0.2297
BL	82.267	0.2126	0.3233	68.150	-0.482	0.267	60.120	-0.463	0.397
BR	91.667	-0.1431	0.2412	81.630	-0.677	0.634	74.020	-0.625	0.416
CL	95.954	-0.1291	0.2633						
CR	97.317	-0.1231	0.1287						
DL	91.326	-0.1580	0.1955	47.760	-0.993	0.430	60.180	-0.929	0.365
DR	90.025	-0.2263	0.3272						
EL	89.189	-0.1881	0.3126						
ER	85.452	-0.1713	0.3045	57.450	-0.386	0.497	71.850	-0.426	0.314
FL	95.773	-0.0967	0.2897						
FR	94.372	-0.1590	0.2273	69.780	-0.649	0.356	72.620	-0.411	0.272
GL	94.030	-0.1620	0.2421						
GR	86.494	-0.1374	0.3471	76.260	-0.558	0.691	67.850	-0.189	0.322
Avg	90.59	-0.12	0.27	66.91	-0.53	0.50	68.70	-0.53	0.32
Stdev	4.75	0.10	0.06	10.51	0.25	0.17	5.63	0.24	0.07

*Only n=8 as the other n=6 were not extracted from the tooth socket

